	Application No.	Applicant(s)
Notice of Allowability .	10/766,369	HORNER ET AL.
	Examiner	Art Unit
	Brian Szmal	3736
The MAILING DATE of this communication apperature All claims being allowable, PROSECUTION ON THE MERITS IS herewith (or previously mailed), a Notice of Allowance (PTOL-85) NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RI of the Office or upon petition by the applicant. See 37 CFR 1.313	(OR REMAINS) CLOSED in or other appropriate communication is su	this application. If not included nication will be mailed in due course. THIS
1. This communication is responsive to the amendment and remarks filed on August 5, 2005.		
2. The allowed claim(s) is/are <u>1 and 3-28</u> .		
3. The drawings filed on <u>28 January 2004</u> are accepted by the Examiner.		
4.		
 Attachment(s) 1. ☐ Notice of References Cited (PTO-892) 2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948) 3. ☑ Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date 8-11-05 4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material 	6. ⊠ Interview Su Paper No./N 98), 7. ⊡ Examiner's A	ormal Patent Application (PTO-152) mmary (PTO-413), Mail Date Amendment/Comment Statement of Reasons for Allowance

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Paul Evans on August 23, 2005.

The application has been amended as follows: Cancel Claim 2.

2. The following is an examiner's statement of reasons for allowance: After reviewing the amendment and remarks filed on August 5, 2005, the claims were found to be allowable since no prior art could be found teaching or suggesting a medical needle shield apparatus comprising: a binding member further comprising a retainer extending therefrom such that the retainer is engageable with the needle cannula to prevent inclination of the binding member while the retainer is engaged with the needle cannula, as claimed in Claims 1, 23 and 28.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Art Unit: 3736

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Brian Szmal whose telephone number is (571) 272-4733. The examiner can normally be reached on Monday-Friday, with second Fridays off.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Max Hindenburg can be reached on (571) 272-4726. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

BS

MAX F. HINDENBUM SUPERVICE THE EXAMINER